


Digital PCR Seminar

Lou Ann Bierwert – ITI Center for Molecular Biology

dPCR involves performing PCR with end-point data collection in a large number of separate reaction chambers, or partitions. Results are obtained by counting the number of partitions in which the amplified target sequence is detected (regarded as positive) and the number of partitions in which there is no amplification (regarded as negative). Absolute quantification of the mean number of target sequences per partition is achieved by applying a Poisson correction to the fraction of the positive partitions. This compensates for the fact that more than one copy of template may be present in some partitions.

“Absolute quantification” used in dPCR refers to an estimate derived from the count of the proportion of positive partitions relative to the total number of partitions and their known volume.



Unlike qPCR, in which the quantification cycle (C_q) depends on variable features such as the instrument, fluorescent reporter dye, and assay efficiency, dPCR relies on a simple count of the number of successful amplification reactions. The counting of positive partitions in an ideal dPCR is definitive and does not require a calibration curve to convert C_q to copy number; knowing the partition number and volume is sufficient.

The ability to measure extremely low concentrations of specific DNA sequences, independent of a standard curve, with high precision, in a complex background, is unique to dPCR.

dPCR was initially developed to investigate minority target measurement, for which rare variants are measured in the presence of large numbers of wild-type sequences. Detection and quantification of rare mutations can provide a useful tool in several scenarios such as the diagnosis and staging of cancer.

Digital PCR has many applications, including the detection and quantification of:

- low-level pathogens**
- rare genetic sequences**
- copy number variations associated with chromosomal rearrangements or gene/chromosomal dosage**
- relative gene expression in tissues or single cells**
- cDNA concentrations**
- New Digital PCR Application for Allele-Specific Copy Number Analysis**

Experimental Workflow on the QuantStudio™ 3D 3D Digital PCR System


- Instrument reads one chip at a time
- Less than a minute to read one chip
- Factory calibrated to detect FAM™, VIC® and ROX® dyes; also compatible with SYBR Green I assays (detects SYBR on the FAM channel and can detect HEX on the VIC channel). ROX is the reference dye in MM so can only singleplex or duplex.
- Intuitive touch screen operation
- Upload results to cloud for analysis
- Instrument stores results for last ~600 chips (max)



- 20,000 reaction wells per chip
 - Minimal sample loss
 - One sample per chip
 - Simple and consistent loading
 - Sealed consumable minimizes contamination
 - Each chip identified by unique 2D barcode
 - Fixed reaction volume minimizes upfront sample manipulation
-



- Total reaction volume loaded on chip is $\sim 14.5 \mu\text{l}$
- Reaction volume is an important determinant of sensitivity.
- Partition volume is 755 pL
- Chip surface has hydrophobic coating to enable isolation of independent reactions.



•dPCR only measures the copies that **are amplifiable**.

•As such there are sequence and sample specific factors to consider prior to your experiment

Factors that affect template amplification

- Sequence damage or DNA integrity
- Assay inhibition or poor sensitivity of assay (primer design)
- Molecular dropout due to linked targets or secondary structure (RE can help, multiple primer design may be necessary)
- Chemical modifications (e.g. formalin crosslinking)
- Denaturation state (single vs double stranded)
- For RNA templates, reverse transcription efficiency

Preparing DNA Samples

- DNA Quality
 - Use an optimized DNA extraction protocol, preferably ending in water
 - salting-out procedures and crude lysates are not recommended
 - Make sure DNA extracts do not contain PCR inhibitors
 - $A_{260/230}$ and $A_{260/280}$ ratios should be between 1.7 and 1.9
 - ~2.0 for RNA
 - Make sure DNA is not degraded
 - E.g. as visualized on an agarose gel
-

- DNA Quantity
- Thermo recommends the following methods of quantitation:
 - nucleic acid quantitation using the Qubit® 3.0
 - Spectrophotometry (OD260) for quality ratios
- The volume of sample added to a digital PCR reaction depends on the
 - Concentration of genomic or complementary DNA (gDNA or cDNA) present in each sample
 - Number of copies of the target sequence present in the genome or total RNA of your samples.
- If target copy number is unknown, qPCR data can be used if available.
- Note-assays should be tested with regular or qPCR to verify if possible.

- Goal is to dilute the samples so that each partition will contain, on average, **~0.6 to ~1.59** copies of the target sequence.

- **Example for human gDNA templates**

- Human genomic DNA has 3.3 pg/copy of a given gene (*E. coli* 0.004 pg/copy)
 - Each partition is 755 pL
 - To determine an appropriate copy number per chip:
-

- **Low end of range:**

0.6 copies / 755 pL x 20,000 partitions = 12,000 copies total->
~795 copies/ μ L or 2.62 ng/ μ L in the dPCR reaction mix.

- **High end of range:** 1.59 copies / 755 pL x 20,000 partitions = 31,800 copies total needed->
~2105 copies/ μ L or 6.94 ng/ μ L in the dPCR master mix.

(I don't follow these numbers – easier to calculate total DNA needed – 31,800 copies * 3.3pg/copy = 104.9ng in 14.5 μ l so 104.5/29 half μ ls = 3.62 so add this to total needed = 104.9 + 3.62 = 108.5ng total in 15 μ l)

**Highest precision (meaning Poisson stats calculations) is achieved at 1.59 copies per partition*

Prepare the digital PCR reactions


- **Required items**

- TaqMan® Assay(s) (\$140-170/200 rxns)
- QuantStudio™ 3D Digital PCR Master Mix (\$140/200 rxns)
- Pipettes and tips, P10 to P1000
- Reaction tubes
- Molecular grade water
- Microcentrifuge
- Vortex
- Gloves, marker pen, lint-free wipes
- Chips (comes with all loading consumables) - \$52/12 (\$884/200)
- \$5-6 per sample

Material	Volume (µL)	Stock	Final
QuantStudio™ 3D Digital PCR Master Mix, 2X	7.5	2X	1X
TaqMan® Assay, 20X (primer/probe mix)	0.75	20X	1X
Diluted DNA	1.5	23 ng/µL*	2.3 ng/µL*
Water	5.25	-	-
Total volume (sample/1 chip)	15	-	-

* Just an example, it will depend on the application

QuantStudio™ 3D Workflow with Chip Loader






Wait for chips to reach room temperature.

- Set the destination for the imaging data
- You may, for example, use a USB memory stick to collect the run data

- Open the chip tray and load the Digital PCR 20K Chip face-up into the bay. Confirm that the Digital PCR 20K Chip is correctly aligned within the chip tray, then close it.



- 1 Orient the chip ID and fill port toward the front of the instrument.
- 2 Confirm that the chip window is clean and correctly aligned to the chip.



Enter a prefix or regret will ensue. You have about 20 seconds.

- FAM™ and VIC® data will be shown after analysis is complete
- After reviewing the results of the run, touch Done
- Further analysis using cloud- or server-based AnalysisSuite software is recommended

- You may run the imaged Digital PCR 20K Chip again **for up to 1 hour after thermal cycling.**
- If you read multiple Chips in rapid succession, touch the scroll buttons to review the results of the previously imaged chips.
- The Instrument retains a copy of the analyzed data for the imaged Chip that you can access from the Run History screen



SETTINGS

Instrument Settings

Data Destinations

Maintenance & Service

Run History

Note: The results will remain in the QuantStudio™ 3D Instrument cache for up to ~600 readings. After ~600 chips, the instrument removes the oldest data file in the cache to store each new reading

Analyze your data in AnalysisSuite in the cloud


- URL: <https://www.thermofisher.com/ca/en/home/life-science/pcr/digital-pcr/quantstudio-3d-digital-pcr-system.html>
- Alternatively, just go to Thermofisher.com and search “AnalysisSuite”
- All users **must create a new login**, or use a pre- existing (validated) account username and password (previous Life Technologies accounts would have been transferred to Thermofisher.com and can still be used)
- Case-sensitive
- Make sure to use **Google Chrome**
- Once logged in, you are taken to this page where you can look at and add to old projects or create a new one.

<input type="checkbox"/>	Project	# of Imported Chips	Last modified
<input type="checkbox"/>	Carlsbad Training	7	Dec 31, 2015 12:03 PM
<input type="checkbox"/>	50 100 ng 2015-12-3	3	Dec 04, 2015 02:28 PM
<input type="checkbox"/>	Cre LoxP project	2	Nov 30, 2015 01:12 PM
<input type="checkbox"/>	CMV Standard curve 3 - chip reread	7	Nov 22, 2015 01:12 AM
<input type="checkbox"/>	RQ Demo Project	5	Nov 12, 2015 03:09 PM
<input type="checkbox"/>	CNV Demo Kit	6	Nov 12, 2015 03:08 PM
<input type="checkbox"/>	SIRS samples	12	Nov 12, 2015 11:13 AM
<input type="checkbox"/>	DilutionCurve	12	Nov 12, 2015 10:48 AM
<input type="checkbox"/>	Illumina Library Quant. Demo Kit	2	Nov 06, 2015 04:39 PM

Either open an existing project, or at bottom of screen, can import a project/create a new one



Project name you created



Define chips

Type into the fields to enter new settings or select the chip(s) and click "Assign settings to multiple chips". NOTE: "Rare Dye" specifies the dye that is used as the numerator in reported ratio results.

What dyes are on your Chip ? VIC FAM

Do you want rare mutation analysis ? Rare Mutation FAM

Actions

- Delete chip(s)
- Import Settings
- Export Settings
- Assign settings to multiple chips

If performing rare allele detection, check this box and designate which reporter

By clicking on "Actions", can import/export 'Define Chips' table, or assign settings to >1 chips simultaneously

Chip	Sample
5A_X210250A_130508_163151.eds	Sample_1
5C_X2102409_130508_180829.eds	Sample_2
6C_X210017C_130508_181054.eds	Sample_3


Y-axis -6000 to 21000
 Point Size 1 - 10
 Point Opacity 0.1 - 1
 Apply axis settings to assay

Review Quality and Calls Double click a chip to view and adjust the quality threshold until you have an acceptable balance of data quantity and quality.

User can adjust plot settings

Actions ▾

5A_X2...63151 Sample_1
 5C_X2...80829 Sample_2
 6C_X2...81054 Sample_3
 7A_X2...65326 Sample_4
 7B_X2...73436 Sample_5



RQ Demo Project Edit project name

Project Dashboard Import Data Define Chips Review Data

Review Quality and Calls Double click a chip to view and adjust the quality th

Chip view
 6C_X210017C_130508_181054.eds
 Sample Name:Sample_3
 Assay Name:assay(default)

Actions ▾
 Color By ▸
 Group By ▸

magnifier

Data points above threshold 17,029 18,488

- Double-clicking chip will expand it
- Default view is to “Color by Call” (i.e. FAM +ve, VIC +ve, FAM & VIC +ve, or No Amp)
- Look for random distribution of positive partitions
- Moving cursor over chip image reveals a “magnifier”
- Clicking “Actions” reveals the ability to “Color by Quality”

17,029 # of wells that passed the quality threshold cutoff


18,488 # of wells total that were filled with dPCR reaction

Project Dashboard Import Data Define Chips Review Data See Results Export

Results

Show settings

Review the results, sort and adjust settings in the table to alter the barchart display as needed. Move the cursor over the image to view more information. Change confidence level, desired precision, and coloring schemes by clicking the "Show settings" button.



For nested sorting, click and drag column headers into this area.

Color	Assay	Target	Sample	Copies/µL	CI Copies/µL	Precision	Chips	Recommendation
Green	FAMVIC	FAM	Sample 1	8.21	6.805 -- 9.906	20.651%	1	
Light Green	FAMVIC	VIC	Sample 1	121.25	115.34 -- 127.45	5.118%	1	
Yellow-Green	FAMVIC	FAM	Sample 2	89.015	83.977 -- 94.356	6%	1	
Yellow	FAMVIC	VIC	Sample 2	7.61E-2	1.07E-2 -- 0.54	609.93%	1	
Orange	FAMVIC	FAM	Sample 3	4.487	3.491 -- 5.767	28.525%	1	
Light Orange	FAMVIC	VIC	Sample 3	128.12	122.11 -- 134.43	4.926%	1	
Blue	FAMVIC	FAM	Sample 4	4.915	3.847 -- 6.279	27.762%	1	
Light Blue	FAMVIC	VIC	Sample 4	102.25	96.002 -- 108.4	5.821%	1	

As the sample becomes more concentrated, the chance of more than 1 molecule being present within a positive partition increases. However, the distribution of molecules throughout the partitions approximates a Poisson distribution and a Poisson correction is applied.

Note how precision skyrockets with few data points. Can't be helped with rare events.

Filter can be applied to any column with a funnel icon.

NHAR Sperm 060116

[Edit project name](#)

[Add comment](#)


[Project Dashboard](#)
[Import Data](#)
[Define Chips](#)
[Review Data](#)
[See Results](#)
[Export](#)

Color by: Sample Assay Target User defined
 Confidence level (%)
 Desired precision (%)
 Quantification Algorithm :

[Hide settings](#)

Results

Review the results, sort and adjust settings in the table to alter the barchart display as needed. Move the cursor over the image to view more information. Change confidence level, desired precision, and coloring schemes by clicking the "Show settings" button.



Under “Show Settings”, you can change the confidence level (90, 95, or 99% - error bars grow with higher confidence desired), precision desired (wide range, 10% standard), and algorithm used (Poisson or Poisson plus).

Export

Export creates a .csv file. All data are exported in a fixed order and format regardless of the sorts and column organization in the tables below.

- Project data as CSV  Export data as .csv, further analysis in 3rd party graphing program
 - Chip data as XML  Export well fluorescence values (for advanced users)
-

- All data are exported in a fixed format that does not reflect column sorting performed in AnalysisSuite™ prior to export.

Project name	test	Assays	1
User name	lbierwer@smith.edu	Samples	1
Created by	lbierwer@smith.edu	Chips	1
Created on	Jun 03, 2016 10:48	Confidence level	95 %
Last modified on	Jul 05, 2016 15:17	Desired precision	10 %
Application version	3.0.3	Algorithm version	0.22
Project description		Algorithm type	Poisson

Rare Target Results

Assay	Sample	Target/Total	CI Target/Total	Copies/μL (VIC)	CI Copies/μL (VIC)	Precision (VIC)	Copies/μL (FAM)	CI Copies/μL (FAM)	Precision (FAM)	Chips	Recommendation
FAMVIC	Sample_1	33.631%	25.414% -- 44.457%	10.333	8.701 -- 12.271	18.756%	5.236	4.113 -- 6.664	27.285%	1	

Quantification Results

Assay	Target	Sample	Copies/μL	CI Copies/μL	Precision	Chips	Recommendation
FAMVIC	FAM	Sample_1	5.236	4.113 -- 6.664	27.285%	1	
FAMVIC	VIC	Sample_1	10.333	8.701 -- 12.271	18.756%	1	

Replicates

Assay	Sample	Dilution	Chip	Target/Total	CI Target/Total	Copies/Rxn (VIC)	CI Copies/Rxn (VIC)	Copies/Rxn (FAM)	CI Copies/Rxn (FAM)	Copies/μL (VIC)	CI Copies/μL (VIC)
FAMVIC	Sample_1	1	test_160603_104616_C02OS8.eds	33.631%	25.414% -- 44.457%	7.80E-3	6.57E-3 -- 9.26E-3	3.95E-3	3.11E-3 -- 5.03E-3	10.333	8.701 -- 12.271

Digital calls

Chip	Dye	# of Neg	# qualified by QT	Threshold
test_160603_104616_C02OS8.eds	FAM	16663	16729	1700.03
test_160603_104616_C02OS8.eds	VIC	16599	16729	2073.29

Home Layout Tables Charts SmartArt Formulas Data Review

Font: Calibri (Body) 12

Alignment: abc, Wrap Text, Merge

Number: General

Format: Conditional Formatting, Styles

Cells: Insert, Delete, Format

Themes: Aa

A78		fx													
	A	B	C	D	E	F	G	H	I	J	K	L	M	N	
1	Project Name	NHAR Sperm 060116	Assays												
2	User Name	lberwer@smith.edu	Samples	16											
3	Created by	lberwer@smith.edu	Chips	16											
4	Created on	6/1/16 17:35	Confidence Level	95%											
5	Last modified on	7/25/16 11:09	Desired Precision	10%											
6	Application version	3.0.3	Algorithm version	0.22											
7	Project Description		Algorithm Type	Poisson											
8	[Quantification Results]														
9	Assay	Target	Sample	Copies/microliter	CI Copies/microliter	Precision	Chips	Recommendation							
10	FAMVIC	FAM	ClD	471.7	459.05 - 484.71	2.76%	1								
11	FAMVIC	FAM	ClU	454.76	442.38 - 467.49	4.82%	1								
12	FAMVIC	FAM	ClI	159.36	152.61 - 166.4	4.82%	1								
13	FAMVIC	FAM	ClV	0.584	0.292 - 1.167	99.96%	1								
14	FAMVIC	FAM	ClZ	2.101	1.37 - 3.222	53.37%	1								
15	FAMVIC	FAM	ClY	116.2	109.37 - 123.23	6.85%	1								
16	FAMVIC	FAM	ClX	2.498	1.785 - 3.496	39.95%	1								
17	FAMVIC	FAM	ClS	327.36	317.29 - 337.76	3.18%	1								
18	FAMVIC	FAM	ClU	199.86	192.04 - 207.99	4.07%	1								
19	FAMVIC	FAM	ClV	7.67E-02	1.08E-2 - 0.544	609.93%	1								
20	FAMVIC	FAM	ClW	0.291	0.109 - 0.775	166.45%	1								
21	FAMVIC	FAM	ClT	0.218	7.04E-2 - 0.677	210.06%	1								
22	FAMVIC	FAM	Sample 1	8.21	6.805 - 9.905	20.65%	1								
23	FAMVIC	FAM	Sample 2	121.25	115.34 - 127.45	5.12%	1								
24	FAMVIC	FAM	Sample 3	89.015	83.977 - 94.356	6.00%	1								
25	FAMVIC	FAM	Sample 4	7.61E-02	1.07E-2 - 0.54	609.93%	1								
26	FAMVIC	FAM	Sample 5	4.487	3.491 - 5.767	28.52%	1								
27	FAMVIC	FAM	Sample 6	128.12	122.11 - 134.43	4.93%	1								
28	FAMVIC	FAM	Sample 7	4.915	3.847 - 6.279	27.76%	1								
29	FAMVIC	FAM	Sample 8	102.35	96.902 - 108.1	5.62%	1								
30	FAMVIC	FAM	Sample 9	221.94	213.71 - 230.49	3.85%	1								
31	FAMVIC	FAM	Sample 10	1.593	1.038 - 2.443	53.37%	1								
32	FAMVIC	FAM	Sample 11	149.88	143.26 - 156.81	4.62%	1								
33	FAMVIC	FAM	Sample 12	7.53E-02	1.06E-2 - 0.535	609.93%	1								
34	FAMVIC	FAM	Sample 13	49.944	46.27 - 53.91	7.94%	1								
35	FAMVIC	FAM	Sample 14	0.298	0.112 - 0.794	166.45%	1								
36	FAMVIC	FAM	Sample 15	267.89	258.19 - 277.96	3.76%	1								
37	FAMVIC	FAM	Sample 16	8.56E-02	1.21E-2 - 0.608	609.93%	1								
38	FAMVIC	FAM	Sample 17	5.216	4.12 - 6.604	26.61%	1								
39	FAMVIC	FAM	Sample 18	7.54E-02	1.06E-2 - 0.536	609.93%	1								
40	FAMVIC	FAM	Sample 19	411.81	400.08 - 423.88	2.93%	1								
41	FAMVIC	FAM	Sample 20	7.63E-02	1.07E-2 - 0.542	609.93%	1								
42	[Replicates]														
43	Assay	Target	Sample	Dilution	Chip	Copies/Rxn	CI Copies/Rxn	CI Copies/microliter	CI Copies/microliter	# of Neg	# qualified by QT	# of Filled	Run Date	Comment	
44	FAMVIC	FAM	ClD		1	ct10_160601_165350_C034W5.eds	0.356	0.347 - 0.366	471.7	459.05 - 484.71	12272	17522	18197	6/1/16 16:53	
45	FAMVIC	FAM	ClU		1	ct10_160601_165350_C034W5.eds	0.343	0.334 - 0.353	454.76	442.38 - 467.49	12430	17522	18197	6/1/16 16:53	
46	FAMVIC	FAM	ClI		1	ct11_160601_165459_C037C.eds	0.12	0.115 - 0.126	159.36	152.61 - 166.4	16097	18155	18830	6/1/16 16:54	
47	FAMVIC	FAM	ClV		1	ct11_160601_165459_C037C.eds	4.41E-04	2.20E-4 - 8.81E-4	0.584	0.292 - 1.167	18147	18155	18830	6/1/16 16:54	
48	FAMVIC	FAM	ClZ		1	ct12_160601_165628_C02M7C.eds	1.59E-03	1.03E-3 - 2.43E-3	2.101	1.37 - 3.222	13230	13251	17625	6/1/16 16:56	
49	FAMVIC	FAM	ClY		1	ct12_160601_165628_C02M7C.eds	8.77E-02	8.27E-2 - 9.30E-2	116.2	109.37 - 123.23	12138	13251	17625	6/1/16 16:56	
50	FAMVIC	FAM	ClX		1	ct13_160601_165716_C02N8K.eds	1.89E-03	1.35E-3 - 2.64E-3	2.498	1.785 - 3.496	18013	18047	18806	6/1/16 16:57	
51	FAMVIC	FAM	ClS		1	ct13_160601_165716_C02N8K.eds	0.247	0.24 - 0.255	327.36	317.29 - 337.76	14095	18047	18806	6/1/16 16:57	
52	FAMVIC	FAM	ClU		1	ct15_160601_165800_C02O69.eds	0.151	0.145 - 0.157	199.86	192.04 - 207.99	14852	17271	18021	6/1/16 16:58	
53	FAMVIC	FAM	ClV		1	ct15_160601_165800_C02O69.eds	5.99E-05	6.16E-6 - 4.11E-4	7.67E-02	1.08E-2 - 0.544	17270	17271	18021	6/1/16 16:58	
54	FAMVIC	FAM	ClW		1	ct1NT_160601_165245_C02S22.eds	2.20E-04	8.24E-5 - 5.85E-4	0.291	0.109 - 0.775	18208	18212	18853	6/1/16 16:52	
55	FAMVIC	FAM	ClT		1	ct1NT_160601_165245_C02S22.eds	1.65E-04	5.31E-5 - 5.11E-4	0.218	7.04E-2 - 0.677	18209	18212	18853	6/1/16 16:52	
56	FAMVIC	FAM	Sample 1		1	1_160601_165849_C02W11.eds	6.20E-03	5.14E-3 - 7.48E-3	8.21	6.805 - 9.905	17530	17639	19007	6/1/16 16:58	
57	FAMVIC	FAM	Sample 2		1	2_160601_165930_C034WH.eds	9.20E-02	9.71E-2 - 9.52E-2	121.25	115.34 - 127.45	16096	17639	19007	6/1/16 16:58	
58	FAMVIC	FAM	Sample 3		1	3_160601_165930_C034WH.eds	6.72E-02	6.34E-2 - 7.12E-2	89.015	83.977 - 94.356	16284	17416	18291	6/1/16 16:59	
59	FAMVIC	FAM	Sample 4		1	4_160601_165930_C034WH.eds	5.74E-05	8.09E-6 - 4.08E-4	7.61E-02	1.07E-2 - 0.54	17415	17416	18291	6/1/16 16:59	
60	FAMVIC	FAM	Sample 5		1	5_160601_170022_C031V5.eds	3.39E-03	2.64E-3 - 4.35E-3	4.487	3.491 - 5.767	17976	18037	18809	6/1/16 17:00	
61	FAMVIC	FAM	Sample 6		1	6_160601_170408_C02OR9.eds	9.20E-02	9.22E-2 - 0.101	128.12	122.11 - 134.43	16374	18037	18809	6/1/16 17:00	
62	FAMVIC	FAM	Sample 7		1	7_160601_170118_C03E15.eds	3.71E-03	2.90E-3 - 4.74E-3	4.915	3.847 - 6.279	17216	17280	17887	6/1/16 17:01	
63	FAMVIC	FAM	Sample 8		1	8_160601_170118_C03E15.eds	7.73E-02	7.32E-2 - 8.16E-2	102.35	96.902 - 108.1	15995	17280	17887	6/1/16 17:01	
64	FAMVIC	FAM	Sample 9		1	9_160601_170233_C039VA.eds	0.168	0.161 - 0.174	221.94	213.71 - 230.49	14779	17475	18349	6/1/16 17:02	
65	FAMVIC	FAM	Sample 10		1	10_160601_170233_C039VA.eds	1.20E-05	7.84E-6 - 1.84E-3	1.593	1.038 - 2.443	17454	17475	18349	6/1/16 17:02	
66	FAMVIC	FAM	Sample 11		1	6_160601_170323_C02SA3.eds	0.113	0.108 - 0.118	149.88	143.26 - 156.81	15708	17590	18427	6/1/16 17:03	
67	FAMVIC	FAM	Sample 12		1	6_160601_170323_C02SA3.eds	5.69E-05	8.01E-6 - 4.04E-4	7.53E-02	1.06E-2 - 0.535	17589	17590	18427	6/1/16 17:03	
68	FAMVIC	FAM	Sample 13		1	7_160601_170408_C02OR9.eds	3.77E-02	3.49E-2 - 4.07E-2	49.944	46.27 - 53.91	17123	17781	18802	6/1/16 17:04	
69	FAMVIC	FAM	Sample 14		1	8_160601_170454_C03717.eds	0.298	0.292 - 0.304	267.89	258.19 - 277.96	16238	15471	19208	6/1/16 17:04	
70	FAMVIC	FAM	Sample 15		1	8_160601_170454_C03717.eds	0.202	0.195 - 0.21	8.56E-02	1.21E-2 - 0.608	15470	15471	19208	6/1/16 17:04	
71	FAMVIC	FAM	Sample 16		1	9_160601_170546_C02O58.eds	3.94E-03	3.11E-3 - 4.99E-3	5.216	4.12 - 6.604	17487	17556	18234	6/1/16 17:05	
72	FAMVIC	FAM	Sample 17		1	9_160601_170546_C02O58.eds	5.75E-05	8.02E-6 - 4.04E-4	7.54E-02	1.06E-2 - 0.536	17555	17556	18234	6/1/16 17:05	
73	FAMVIC	FAM	Sample 18		1	10_160601_170642_C02O76.eds	0.311	0.302 - 0.32	411.81	400.08 - 423.88	17271	17360	18260	6/1/16 17:06	
74	FAMVIC	FAM	Sample 19		1	10_160601_170642_C02O76.eds	5.76E-05	8.11E-6 - 4.09E-4	7.63E-02	1.07E-2 - 0.542	17359	17360	18260	6/1/16 17:06	
75	[Digital Call]														
76	Chip	Dye	# of Neg	# qualified by QT	Threshold										
77	10_160601_170642_C02O76.eds	FAM	17721	17360	4.179.56										
78	10_160601_170642_C02O76.eds	FAM	17359	17360	3011.31										
79	1_160601_165849_C02W11.eds	FAM	17530	17639	1690.5										
80	1_160601_165849_C02W11.eds	FAM	16096	17639	1076.61										
81	2_160601_165930_C034WH.eds	FAM	16284	17416	4422.15										
82	2_160601_165930_C034WH.eds	FAM	17415	17416	3233.23										
83	3_160601_170022_C031V5.eds	FAM	17976	18037	2163.17	</									